



Fiber Receiver to BNC (FRB1) BNC to Fiber Transmitter (FTB1)

FRB1/FTB1 Quick Start Guide

Thank you for purchasing an FRB/FTB single channel unit. EHT's fiber products are robust and reliable; with a few simple precautions they will prove to be tremendously useful for your laboratory/industrial control and signaling needs. Should an FRB/FTB unit fail due to a manufacturing defect, we will replace it up to one year from the date of purchase. For questions about operation or troubleshooting assistance, please contact support@eagleharbortech.com.



Precautions

1. The FTB1 BNC to Fiber Transmitter unit can be driven with a maximum voltage of 5.5 V DC at the input; it is internally 50 Ω terminated.
2. The FRB1 Fiber Receiver to BNC unit must be terminated into 50 Ω for clean high speed TTL signals.
3. When preparing 2.2 mm jacketed plastic fiber (we use ESKA SH-4001), be certain to make the end cuts with a razor knife. This provides a much cleaner cut than using cutters, scissors, etc.
4. Be sure that the fiber is fully seated into the barrel of the receiver or transmitter. Loosen the ferrule, insert the fiber until it bottoms out in the receiver or transmitter body, then finger tighten the ferrule. Overtightening is not required as the grip strength of the ferrule on the fiber is significant.

sky blue
microsystems

Sky Blue Microsystems GmbH
Geisenhausenerstr. 18
81379 Munich, Germany
+49 89 780 2970, info@skyblue.de
www.skyblue.de

ZERIF
TECHNOLOGIES LTD.
A SKY BLUE COMPANY FOUNDED 1999

In Great Britain:
Zerif Technologies Ltd.
Woodberry House, 2 Woodberry Grove
Finchley, London N12 0DR
+44 115 855 7883, info@zerif.co.uk
www.zerif.co.uk